

# seca 210

 **Evocare**<sup>®</sup>  
[www.evocare.com.au](http://www.evocare.com.au)  
Ph: 07 3355 8000 Fax: 07 3355 5043  
Email: [sales@evocare.com.au](mailto:sales@evocare.com.au)

**seca**<sup>®</sup>  
Precision for health

## Congratulations!

In the **seca 210** measuring mat you have acquired an accurate and at the same time sturdy piece of equipment for measuring the length of babies.

The **seca 210** measuring mat is designed for use in paediatric practices, paediatric hospitals or for home visits by midwives and paediatricians.

The measuring mat can be rolled up and so requires only a little space. There are also holes in the head stop on the mat to enable it to be hung on a wall.

## Safety

Before using the new measuring mat, please take a little time to read the following safety instructions.

Follow the operating instructions.

- Never leave a baby lying on the measuring mat unsupervised.
- Only measure length on surfaces from which the baby cannot fall.
- Ensure that the measuring mat is cleaned and disinfected regularly to avoid infectious diseases being passed on.
- Ideally, a second person is required to measure length exactly.


## Operation

### Measurement method:

1. Lie the baby on the measuring mat. The baby's head must lie against the head stop (1) of the measuring mat.
2. Using your right hand, push the foot stop (2) in the direction of the baby's feet. The left hand holds the baby in position.

3. Stretch the baby's legs by pressing its knees carefully and gently down to the mat with the left hand.
4. With your right hand, rest the foot stop (2) at right-angles to the soles of the baby's feet.
5. Read off body length at the marking on the measuring mat (3).

If the above steps are followed, accuracy greater than  $\pm 5$  mm is achieved.



## Cleaning

Clean the measuring mat as required using a domestic cleaning agent or commercially available disinfectant. Follow the manufacturer's instructions.

## Technical data

Measurement range: 100 - 990 mm /  
4 - 39 inches  
Graduations: 5 mm / 1/4 inch  
Accuracy: greater than  $\pm 5$  mm  
Dimensions  
(W x H x D): 300 mm x 140 mm x  
1340 mm  
Weight: approx. 500 g  
Temperature range: +10 °C to +40 °C  
Medical device in  
accordance with  
Directive  
93/42/EEC: Class I with  
measuring function

## Disposal

Should it no longer be possible to use the measuring mat, the responsible waste disposal association will be glad to give you information on how to dispose of it properly.

## Warranty

A two-year warranty from date of delivery applies to defects attributable to poor materials or workmanship. All moveable parts - batteries, cables, mains units, rechargeable batteries etc. - are excluded. Defects which come under warranty will be made good for the customer at no charge on production of the receipt. No further claims can be entertained. The costs of transport in both directions will be borne by the customer should the equipment be located anywhere other than the customer's premises. In the event of transport damage, claims under warranty can be honoured only if the complete original packaging was used for any transport and the merchandise secured and attached in that packaging just as it was when originally packed. All the packaging should therefore be retained.

A claim under warranty will not be honoured if the equipment is opened by persons not expressly authorised by seca to do so. We would ask our customers abroad to contact their local sales agent in the event of a warranty matter.



## Declaration of conformity

Mechanical equipment for measuring length  
Model seca 210

Batch no.: see measuring mat

The measuring mat meets the applicable requirements of Directive 93/42/EEC concerning medical devices.

Hamburg, August 2014

Frederik Vogel  
– CEO Development and Manufacturing –  
seca gmbh & co. kg  
Hammer Steindamm 9-25  
22089 Hamburg  
Germany  
Phone: +49 (0)40 20 00 00 0  
Fax: +49 (0)40 20 00 00 50  
①: [www.seca.com](http://www.seca.com)

## EVOCARE AUSTRALIA PTY LIMITED

A.B.N. 98 078 566 604

Trading as **EVOCARE** and **L&M EQUIPMENT**

P.O. Box 1144, Stafford Qld. 4053

Ph: 07 3355 8000 Fax: 07 3355 5043

Website: <http://www.evocare.com.au>

Email: [sales@evocare.com.au](mailto:sales@evocare.com.au)

[workshop@evocare.com.au](mailto:workshop@evocare.com.au)

[warehouse@evocare.com.au](mailto:warehouse@evocare.com.au)

[accounts@evocare.com.au](mailto:accounts@evocare.com.au)

